

香港特別行政區政府
衛生署

香港灣仔皇后大道東 183 號
合和中心 31 樓 3101 室



THE GOVERNMENT OF THE HONG KONG
SPECIAL ADMINISTRATIVE REGION
DEPARTMENT OF HEALTH
ROOM 3101, 31/F, HOPEWELL CENTRE,
183 QUEEN'S ROAD EAST, WANCHAI,
HONG KONG.

本處檔號 OUR REF.: L/M(769) to DH/MDCO/20-10/8
來函檔號 YOUR REF.:
電話 TEL.: 3107 8484
電郵 EMAIL: mdco@dh.gov.hk
圖文傳真 FAX: 3157 1286

Dear Doctor,

FDA Issued Safety Communication on Preventing Surgical Fires

Your attention is drawn to a notice issued by U.S. Food and Drug Administration (FDA) regarding the prevention of surgical fires. The FDA is advising health care professionals, health care facility executives and staff on factors that increase the risk of surgical fires, particularly the delivery of supplemental oxygen to patients.

Surgical fires can occur at any time all three elements of the fire triangle are present:

- Ignition source (e.g., ESUs, lasers, and fiber optic light sources)
- Fuel source (e.g., surgical drapes, alcohol-based skin preparation agents, the patient)
- Oxidizer (e.g., oxygen, nitrous oxide, room air)

In order to minimize the risk of surgical fire, FDA recommends the following:

1. Conduct a fire risk assessment at the beginning of each procedure.
2. Use supplemental oxygen safely.
3. Use alcohol-based (flammable) skin preparation agents safely.
4. Use devices and other surgical equipment safely.
5. Encourage communication among members of the surgical team.
6. Plan how to manage a surgical fire.

For more details about the notice, please refer to the following links:

<http://www.fda.gov/MedicalDevices/Safety/AlertsandNotices/ucm275189.htm>

Thank you for your attention.

Yours faithfully,

(Dr Addi Chan)
for Director of Health

竭誠服務 顧客為本 素質為先

We are committed to providing client-oriented service