

衛生署
醫療儀器管制辦公室
香港灣仔皇后大道東 183 號
合和中心 31 樓 3101 室



DEPARTMENT OF HEALTH
Medical Device Control Office
Room 3101, 31/F, Hopewell Centre,
183 Queen's Road East, Wan Chai, Hong Kong

本署檔號 OUR REF: S/F(194)in DH/MDCO/20-5/7

來函檔號 YOUR REF:

電話 TEL: 3107 8484

電郵 EMAIL: mdco@dh.gov.hk

圖文傳真 FAX: 3157 1286

24 Nov 2010

Dear Healthcare Professionals,

TGA issued alert on the use of Portable Blood Glucose Meters and Strips

Your attention is drawn to an alert issued by the Therapeutic Goods Administration (TGA) of the Australian Government regarding the use of certain blood glucose meters and strips.

Portable blood glucose meters and strips which use the enzyme glucose dehydrogenase pyrroloquinoline quinone (GDH-PQQ), also known as glucose-dye-oxidoreductase, have the potential to detect the sugars galactose and xylose (monosaccharides) and maltose (disaccharide) and hence provide a false high glucose reading in the presence of these substances (e.g. icodextrin-containing peritoneal dialysis solutions, some immunoglobulins, abatacept, etc.). As such, it is advised that patients undergoing treatment with any product that contains or is broken down by the body into maltose, galactose, or xylose should not use blood glucose monitors utilising GDH-PQQ technology. Adverse events have been reported in Australia and worldwide since 2005. Although its overall incidence has not markedly increased, the TGA issued an updated advisory as a precautionary measure. For details of the alert, please refer to the TGA website:

<http://www.tga.gov.au/alerts/devices/blood-glucose-meters.htm>

We understand that portable blood glucose monitors and strips using GDH-PQQ have been distributed in Hong Kong. Please contact your supplier should you have query about the blood glucose monitors you are currently using or recommending to your patients.

Suspected adverse events related to medical devices can be reported to the Medical Device Control Office of the Department of Health (fax: 3157 1286). For details, please refer to the website: <http://www.mdco.gov.hk> under "Reporting Adverse Incidents".

Yours faithfully,

(Dr Addi CHAN)
for Director of Health