

衛生署
醫療儀器管制辦公室

香港灣仔皇后大道東 183 號
合和中心 31 樓 3101 室



DEPARTMENT OF HEALTH
MEDICAL DEVICE CONTROL OFFICE

ROOM 3101, 31/F., HOPEWELL CENTRE,
183 QUEEN'S ROAD EAST, WAN CHAI
HONG KONG

本署檔號 OUR REF. S/F(146) in DH/MDCO/18-25/2

來函檔號 YOUR REF.:

電話 TEL.: 3107 8484

傳真 FAX No.: 3157 1286

1 Aug 2011

Dear healthcare professional,

Recall of Two Intrauterine Contraceptive Devices(IUCDs)
Mona Lisa CuT 380A and NT Cu380

We would like to bring to your attention that the Department of Health (DH) alerted the public on 29 Jul 2011 concerning a recall of all batches of intrauterine contraceptive devices (IUCDs), Mona Lisa CuT 380A and Mona Lisa NT Cu380 (please see attached photos), supplied by CNW Far East Limited (CNWFE), following discovery of mislabeling and wrong packaging involving the two models.

The discovery was made by the DH's Chai Wan Maternal and Child Health Centre (MCHC) when they served a family planning client on 28 Jul 2011. On opening a CuT 380A package, staff discovered that an NT Cu380 device was found instead. CuT 380A has a life span of 10 years while NT Cu 380 has only five years.

DH has contacted the local supplier for information and investigation is in progress. The findings so far revealed that the entire production chain for Mona Lisa IUCD takes place in Belgium.

While DH will be monitoring the recall, you are advised to stop using the products immediately in case you happen to have the affected items in hand. CNWFE has set up a hotline 2388 2933 to answer related enquiries.



Thank you for your attention.

Yours faithfully,

(Dr Addi CHAN)
for Director of Health