

15 May 2012

**Attention: Operating Room Director/ Materials Management**

Dear Customer,

COVIDIEN (formerly United States Surgical, a division of Tyco Healthcare Group, L.P.) is conducting a **VOLUNTARY RECALL** of various production lots of ROTICULATOR™ Single Use Stapler.

This voluntary recall is being conducted due to the potential for the sterility barrier to be compromised. The use of products with this condition may result in a potentially increased risk for infection.

Please note the affected product reference numbers and descriptions are listed below.

<u>REF #</u>	<u>PRODUCT DESCRIPTION</u>
017612	ROTIULATOR™ 55-3.5 Single Use Stapler
017614	ROTIULATOR™ 55-4.8 Single Use Stapler
017615	ROTIULATOR™ 30-3.5 Single Use Stapler
017617	ROTIULATOR™ 30-4.8 Single Use Stapler
017619	ROTIULATOR™ 30-V3 Single Use Stapler

**LOT #'s BEGINNING WITH THE FOLLOWING LETTER/NUMBER COMBINATIONS ARE AFFECTED AND MUST BE RETURNED.**

<b>P7E</b>	<b>P8B</b>	<b>P8L</b>	<b>P9H</b>	<b>P0E</b>	<b>P1B</b>	<b>P1L</b>
<b>P7F</b>	<b>P8C</b>	<b>P8M</b>	<b>P9J</b>	<b>P0F</b>	<b>P1C</b>	<b>P1M</b>
<b>P7G</b>	<b>P8D</b>	<b>P9A</b>	<b>P9K</b>	<b>P0G</b>	<b>P1D</b>	<b>P2A</b>
<b>P7H</b>	<b>P8E</b>	<b>P9B</b>	<b>P9L</b>	<b>P0H</b>	<b>P1E</b>	<b>P2B0153X</b>
<b>P7J</b>	<b>P8F</b>	<b>P9C</b>	<b>P9M</b>	<b>P0J</b>	<b>P1F</b>	
<b>P7K</b>	<b>P8G</b>	<b>P9D</b>	<b>P0A</b>	<b>P0K</b>	<b>P1G</b>	
<b>P7L</b>	<b>P8H</b>	<b>P9E</b>	<b>P0B</b>	<b>P0L</b>	<b>P1H</b>	
<b>P7M</b>	<b>P8J</b>	<b>P9F</b>	<b>P0C</b>	<b>P0M</b>	<b>P1J</b>	
<b>P8A</b>	<b>P8K</b>	<b>P9G</b>	<b>P0D</b>	<b>P1A</b>	<b>P1K</b>	

**Also Best Practice Kits Ref #'s: 00Z0800, 00Z0976, 00Z1401, 00Z1407, 00B0117, 00B0122**

**NOTE: THE SPECIFIC LOT NUMBERS LISTED BELOW AND LOT NUMBERS ENDING IN “”RMX” and “RRMX” ARE NOT AFFECTED BY THIS RECALL AND ARE ACCEPTABLE FOR USE. P2A0194XR, P2B0488XR, P2B0153XR**

**REQUIRED ACTIONS:**


1. Please immediately identify and quarantine all affected inventory.  
**NOTE: Units from the affected lots may have been incorporated into Covidien BEST PRACTICE™ KITS. If you have purchased BEST PRACTICE™ KITS, please evaluate whether these kits contain units from the affected lots. All affected ROTICULATOR™ units must be returned.**
2. Please complete the attached ROTICULATOR™ Single Use Stapler Recalled Product Return form in its entirety. If you do not have any units from the affected lots in your inventory, simply return the product return form indicating you have zero(0) unit and report to our local Covidien office.
3. Please return affected product to its invoicing entity as follows:
  - a. Customers that received product directly from COVIDIEN, please return affected product to our local Covidien office with completed ROTICULATOR™ Single Use Stapler Recalled Product Return form.
  - b. If you purchased product from a Distributor please complete the ROTICULATOR™ Single Use Stapler Recalled Product Return form and contact your Distributor directly. The completed form and all affected units must be returned through the Distributor.

This voluntary recall is limited to the material codes and lot #'s beginning with the letter/number combinations identified and does ***NOT*** affect any other lots of Covidien devices. This voluntary recall is being conducted with the knowledge of the FDA.

We ask that you reply to Covidien **WHETHER OR NOT** you have affected product at your site. Your response is vital to our monitoring of the effectiveness of this recall.

We apologize for any inconvenience this may have caused and thank you for your business and continued support. If you have any questions or concerns, please do not hesitate to contact your Covidien Representative.

Sincerely,



**Mi Youn Chang**  
RA Director - Korea & SEA

**Attachment A: ROTICULATOR™ Single Use Stapler Recalled Product Inventory Form**

Attachment A

**RODICULATOR™ Single Use Stapler Recalled Product Return form**

**Product REF #'s:**

017612	RODICULATOR™ 55-3.5 Single Use Stapler
017614	RODICULATOR™ 55-4.8 Single Use Stapler
017615	RODICULATOR™ 30-3.5 Single Use Stapler
017617	RODICULATOR™ 30-4.8 Single Use Stapler
017619	RODICULATOR™ 30-V3 Single Use Stapler

**LOT #'s BEGINNING WITH THE FOLLOWING LETTER/NUMBER COMBINATIONS ARE AFFECTED AND MUST BE RETURNED.**

<b>P7E</b>	<b>P8B</b>	<b>P8L</b>	<b>P9H</b>	<b>P0E</b>	<b>P1B</b>	<b>P1L</b>
<b>P7F</b>	<b>P8C</b>	<b>P8M</b>	<b>P9J</b>	<b>P0F</b>	<b>P1C</b>	<b>P1M</b>
<b>P7G</b>	<b>P8D</b>	<b>P9A</b>	<b>P9K</b>	<b>P0G</b>	<b>P1D</b>	<b>P2A</b>
<b>P7H</b>	<b>P8E</b>	<b>P9B</b>	<b>P9L</b>	<b>P0H</b>	<b>P1E</b>	<b>P2B0153X</b>
<b>P7J</b>	<b>P8F</b>	<b>P9C</b>	<b>P9M</b>	<b>P0J</b>	<b>P1F</b>	
<b>P7K</b>	<b>P8G</b>	<b>P9D</b>	<b>P0A</b>	<b>P0K</b>	<b>P1G</b>	
<b>P7L</b>	<b>P8H</b>	<b>P9E</b>	<b>P0B</b>	<b>P0L</b>	<b>P1H</b>	
<b>P7M</b>	<b>P8J</b>	<b>P9F</b>	<b>P0C</b>	<b>P0M</b>	<b>P1J</b>	
<b>P8A</b>	<b>P8K</b>	<b>P9G</b>	<b>P0D</b>	<b>P1A</b>	<b>P1K</b>	

**Also Best Practice Kits Ref #'s: 00Z0800, 00Z0976, 00Z1401, 00Z1407, 00B0117, 00B0122**

Date: \_\_\_\_\_

Name of Person Completing form: \_\_\_\_\_ Title: \_\_\_\_\_

Direct Phone # \_\_\_\_\_ Email: \_\_\_\_\_

Account Name: \_\_\_\_\_ Covidien Account number: \_\_\_\_\_

Account address: \_\_\_\_\_

City: \_\_\_\_\_ State: \_\_\_\_\_ Zip code: \_\_\_\_\_

